



Uttarakhand Livestock Development Board

Pashudhan Bhawan, 1st Floor, Mothrowala Road,

P.O. Mothrowala, Dehradun-248001,

Telefax : 0135-2532619

Email: ceo_uldb@rediffmail.com Website: www.uldb.org

GSTIN: 05AABFU8674M1ZK

Letter No. 170/ ULDB/PROC./Vol.XV/2018-19

Dated: 12 April 2019

Clarification & Amendment

Bid Ref. No. 4857/ULDB/PROC./Vol.XV/2018-19

Dated: 08th March 2019

Kindly look in to the following clarification/ amendment in Bid Reference No.4857/ULDB/PROC./Vol.XV/2018-19 Dated: 08th March 2019:

S. No.	Reference	Clarification & Amendment
1	Page No. 31 SCHEDULE: TWELVE PART A TECHNICAL SPECIFICATI ON OF BENCH TOP CENTRIFUGE	<p>Non refrigerated type, Rotor for 5ml with 6 positions or more and 1.5 /2ml with 24 positions with rpm minimum 12000 or more for both the rotors with suitable bio containment lids for both rotors . IEC 61010 ,IVD and European CE (Certified to be provided).Control and Display: For RPM /RCF, Time. LCD Display. Should install the system in the Indian electrical system. Provide 3 KAV or suitable servo stabilizer. Warranty: 2 years.</p> <p>Should be read as</p> <p>Non refrigerated type, Rotor for 5ml with 6 positions or more and 1.5 /2 ml with 24 positions with RCF 21000 or more for both 2 ml and 5 ml rotors with suitable bio containment lids for both rotors. IEC 61010 ,IVD and European CE (Certified to be provided).Control and Display: For RPM /RCF, Time. LCD Display. Should install the system in the Indian electrical system. Provide 3 KAV or suitable servo stabilizer. Warranty: 2 years.</p>

Other Terms & conditions will remain same as described in Bid Reference No. 4857/ULDB/PROC./Vol.XV/2018-19 Dated: 08th March 2019 and 5126/ ULDB/PROC./Vol.XV/2018-19 Dated: 29th March 2019.

Chief Executive Officer

Uttarakhand Livestock Development Board